



Certificate of Analysis

Product No.	42	Product Name	Resibufogenin										
Lot. No.		422411	Unit Quantity	20mg									
Analysis Date		2024-12-06	Storage Conditions	Refrigerate (≤10° C)									
<chem>CC12CCC3C(C1)C(=O)OC3C2C4C(C)CC(O)CC4</chem> C₂₄H₃₂O₄ MW : 384.51 CAS RN: 465-39-4													
This product conforms to the specifications in the 18th edition of the Japanese Pharmacopoeia , Reagents, Test Solutions, " Resibufogenin for assay" and "Resibufogenin for thin-layer chromatography".													
Content (Not less than 98.0%)			Passed (100%)										
 <table><tr><th>Retention time</th><th>Peak area</th><th>%</th></tr><tr><td>9.22</td><td>12001591</td><td>100.00</td></tr><tr><td>Total</td><td>12001591</td><td>100.00</td></tr></table>					Retention time	Peak area	%	9.22	12001591	100.00	Total	12001591	100.00
Retention time	Peak area	%											
9.22	12001591	100.00											
Total	12001591	100.00											
assay (HPLC)	Detector : An ultraviolet absorption photometer (300nm) Column : YMC-Pack ODS-AL (ϕ 4.6 × 250mm) Column temperature : A constant temperature of about 40°C. Mobile phase : A mixture of water and acetonitrile(1 : 1).												

		Flow rate : Adjust so that the retention time of resibufogenin is about 9 minutes.
		Time span of measurement : About twice as long as the retention time of resibufogenin, beginning after the peak of solvent.
Description		Passed
Absorbance $E_{1\text{cm}}^{1\%}$ (300nm) : 131~145		Passed (144)
Purity Related substances (TLC)		Passed
Water (by coulometric titration)		4.9%
Precautions Order quantities that can be used immediately, and after receipt of a reference standard, store it at the specified temperature and use it as soon as possible. The expiration date will be set sequentially based on the results of stability testing.		

IWASHIMA YAKUHI Co., Ltd.

4-201, Sakurazaka, Moriyama-ku, Nagoya City, Aichi Prefecture 463-0018, Japan

TEL : +81-52-715-5601

FAX : +81-52-715-5602