



Certificate of Analysis

Product No.	06	Product Name	Eleutheroside B
Lot. No.	062501	Unit Quantity	20mg
Analysis Date	2025-09-05	Storage Conditions	Room temperature



C₁₇H₂₄O₉
MW : 372.37
CAS RN: 118-34-3

This product conforms to the specifications in the 19th edition of the Japanese Pharmacopoeia, Reagents, Test Solutions, "Eleutheroside B for liquid chromatography".



Retention time	Peak area	%
4.770	2861	0.08
10.187	3740903	99.37
12.550	10570	0.28
13.567	4939	0.13
22.930	5531	0.15
Total	3764804	100.00

Area percentage	: Not less than 98.0%
Detector	: An ultraviolet absorption photometer (265nm)
Column	: Wakosil- II 5C18 HG (4.6×150mm)

Purity Related substances (HPLC)	Column temperature	A constant temperature of about 50°C.	Passed (99.4%)
	Mobile phase	: A mixture of water and acetonitrile(9 : 1).	
	Flow rate	: Adjust so that the retention time of eleutheroside B is about 10 minutes.	
	Time span of measurement	: About 3 times as long as the retention time of eleutheroside B,	
Description			Passed
Identification (UV)			Passed
Water (by coulometric titration)			5.1%
Precautions Order quantities that can be used immediately, and after receipt of a reference standard, store it at the specified temperature and use it as soon as possible.			
Expiration Date		September 2029 (unopened)	

IWASHIMA YAKUHHIN Co., Ltd.

4-201, Sakurazaka, Moriyama-ku, Nagoya City, Aichi Prefecture 463-0018, Japan

TEL : +81-52-715-5601

FAX : +81-52-715-5602